



Addendum 1 – Responses to Questions

RFP#: 24-101A

Issue Date: June 6, 2024

Q1: Do the pellets need to be manufactured aseptically, considering that the final packaged product will undergo terminal sterilization using either the Gamma or E-beam method?

A1: The pellets do not need to be manufactured aseptically, and a terminal sterilization of the final packaged product will suffice.

Q2: Will quotes supporting only the analytical portion of the RFP scope be considered?

A2: CONRAD requires a CDMO who can provide both manufacturing and analysis.

Note: A signed acknowledgement of this addendum must be attached to your proposal. Signature on this addendum does not substitute for signature on the original proposal document. The original proposal document must be signed where indicated.

Name of Firm

Signature/Title

Date